

## ABRAXIS® Spinosyn Coated Tube ELISA 50020B

1.

Label test tubes for Standards (Calibrators), Control, and Samples.

Add 500 μL of either Standards, Control or Samples down the inside wall of each test tube by aiming the pipet tip 1/4" to 1/2" below the tube rim without touching the rim or tube wall with the pipet tip; deliver liquid gently. 6.


7.




Add 4 mL of Washing Solution to each tube (alternatively flood the tubes completely with wash solution then invert to empty tubes). Using a smooth motion, invert tubes over a sink and pour out the tube contents: keep inverted and blot the test tube rims on several layers of paper toweling.

Repeat this step 4 times for a total of five washes.

2.



Add 250 µL of the Spinosyn Antibody Solution to the bottom of each tube by inserting the pipette tip all the way into the bottom of the tube without touching the side of the tubes.



Add 500 μL of Color Reagent down the inside wall of each tube by using the technique described in Box 2.

3.



React 15 minutes at room temperature (15 °- 30°C).

8.




React for 15 minutes at room Temperature (15°-30° C). During this period, add 1 mL of Washing Solution into a clean tube for use as an instrument blank in Step 8.

4.



Add 250 μL of Spinosyn Enzyme Conjugate down the inside wall of each tube by using the technique described in Box 2. Vortex or swirl for 5 to 10 seconds 9.



Add 500 µL of Stopping Solution down the inside wall of each tube by using the technique previously Described. Read results at 450 nm within 15 minutes after adding the Stopping Solution. Multiply results of samples by the appropriate dilution factor (if any).

[Safety Caution: Stopping Solution contains diluted sulfuric acid.]

**5**.



React 15 minutes at room temperature (15 °- 30°C).

Gold Standard Diagnostics 795 Horsham Road Horsham, PA 19044 WEB: www.abraxiskits.com T (215) 357 3911 F (215) 357 5232

Ordering: info.abraxis@us.goldstandarddiagnostics.com Technical Support: support.abraxis@us.goldstandarddiagnostics.com

Version: 01

Date this Flow Chart is effective: 05/17/2024