


# ABRAXIS® Domoic Acid ELISA ON0021


## 1. Addition of Standards, Samples

Add 50 uL of standard solutions, control or samples.




## 2. Addition of Enyzme Conjugate

Add 50 uL of enzyme conjugate.




# 3. Addition of Antibody Solution

Add 50 uL of the antibody solution. Cover and mix for 30 seconds by rotating on benchtop. Incubate for 60 minutes at room temperature.




#### 4. Washing of Plates

Wash the plates three times with 1X washing buffer. Fill to overflow.




# 5. Addition of Color Solution

Add 100 uL of substrate/color solution. Incubate 30 minutes at room temperature and away from direct sunlight.




#### **6.** Addition of Stopping Solution

Add 50 uL of stop solution.



## 7. Measurement of Color

Measure color at 450 nm. Calculate results.



Gold Standard Diagnostics 795 Horsham Road Horsham, PA 19044 WEB: www.abraxiskits.com

Version: 01

Date this Flow Chart is effective: 05/17/2024