

## Glyphosate in Pureed Baby Food Sample Preparation

### 1. Intended Use

For the detection of Glyphosate in pureed baby food products.

### 2. Sensitivity

7.5 ppb in matrix

### 3. Materials and Reagents Required

Analytical balance

Microcentrifuge tubes, 2.0 mL

Microcentrifuge capable of spinning at 8,000 x g or 10,000 rpm

Glass vials – 4 mL and 20 mL with Teflon-lined caps

Micropipettes with disposable plastic tips

Deionized water

ABRAXIS® Glyphosate Clean-up Reagent (PN 500094)

Serological pipettes, 5 mL or 10 mL

Rotator and/or shaker

Vortex mixer

ABRAXIS® Glyphosate Sample Diluent (PN 500082)

ABRAXIS® Glyphosate ELISA kit (PN 500205)

### 4. Notes and Precautions

This procedure is intended for use with pureed baby food products. Other matrices should be thoroughly validated before use with this procedure.

- Analysis should be performed with the ABRAXIS® Glyphosate Plate ELISA Kit as soon as possible after extraction. Samples should not sit more than one day in plastic microcentrifuge tubes before being diluted and analyzed.
- This procedure is for research use only. It is not intended for diagnostic procedures.

### 5. Procedure

5.1 Measure 0.5 g of sample into an appropriately labeled 20 mL glass vial.

5.2 Add 10 mL of deionized water to the sample (1:20 dilution).

5.3 Vortex vigorously for 10 – 15 seconds. Place sample on rotator or shaker for 10 minutes.

5.4 Remove from rotator or shaker and allow the sample to settle for at least 2 minutes.

5.5 Transfer 1.5 mL of the sample to an appropriately labeled 2.0 mL microcentrifuge tube containing 50 mg of ABRAXIS® Glyphosate Clean-up Reagent and vortex for 30 seconds.

5.6 Centrifuge for 5 minutes at ~8,000 x g. Make sure the centrifuge is properly balanced.

5.7 Add 800 µL of ABRAXIS® Glyphosate Sample Diluent to an appropriately labeled 4 mL glass vial. Add 200 µL of the supernatant (from 5.6) to the ABRAXIS® Glyphosate Diluent in the vial (1:5 dilution). Vortex.

5.8 This will then be analyzed as sample, see *Derivatization of Standards, Control and Samples* in the Test Preparation section of the ABRAXIS® Glyphosate Plate ELISA Kit user's guide.

### 6. Evaluation of Results

The ELISA results must be multiplied by a factor of 100 to account for the necessary dilution. Samples showing a concentration lower than Standard 1 (0.075 ppb) should be reported as < 7.5 ppb of Glyphosate. Samples showing a higher concentration than Standard 5 (4.0 ppb) can be reported as > 400 ppb or diluted further and re-analyzed to obtain an accurate quantitative result.

**7. For ordering or technical assistance contact:**

Gold Standard Diagnostics

795 Horsham Road

Horsham, PA 19044

WEB: [www.abraxiskits.com](http://www.abraxiskits.com)

Phone: (215) 357 3911

Fax: (215) 357 5232

Ordering: [info.abraxis@us.goldstandarddiagnostics.com](mailto:info.abraxis@us.goldstandarddiagnostics.com)

Technical Support: [support.abraxis@us.goldstandarddiagnostics.com](mailto:support.abraxis@us.goldstandarddiagnostics.com)

Date this Technical Bulletin is effective: 05/16/2024

Version: 01