

## Glyphosate in Tofu Sample Preparation

### 1. Intended Use

For the detection of Glyphosate in tofu (silken, soft, medium, firm, extra firm).

### 2. Sensitivity

0.075 ppm in matrix

### 3. Materials and Reagents Required

Analytical balance

20 mL glass vials with Teflon-lined caps

Microcentrifuge tubes

4 mL glass vials with Teflon-lined caps

Disposable pipettes

Scoopula

Food processor or resealable plastic bags

Micropipettes with disposable plastic tips

Vortex mixer

Microcentrifuge

Timer

Plate shaker or Micro-well plate holder with insert retainer for vortex mixer

1 N Hydrochloric acid (HCl)

ABRAXIS® Glyphosate Sample Diluent (PN 500082)

ABRAXIS® Glyphosate Plate ELISA Kit (PN 500205)

### 4. Notes and Precautions

This procedure is intended for use with tofu (silken, soft, medium, firm, extra firm). Other matrices should be thoroughly validated before use with this procedure.

- Samples must be thoroughly homogenized before extraction. Samples packed in liquid should be drained prior to homogenizing. Samples can be homogenized in a food processor or by placing in an appropriately labeled resealable plastic bag and kneading thoroughly, until a uniform, smooth consistency is obtained.
- Hydrochloric Acid must be handled with care. Wear appropriate protective clothing (gloves, glasses, etc.). Avoid contact with skin and mucous membranes. If contact occurs, wash with copious amounts of water and seek appropriate medical attention.
- Due to the viscous nature of the sample extracts, the microtiter plate should be placed on a plate shaker or vortex mixer fitted with a micro-well plate holder adapter for the incubations with the antibody and conjugate solutions. This will allow for the appropriate mixing of all reagents in the microtiter wells.

### 5. Procedure

5.1 Weigh 1 g of homogenized sample into an appropriately labeled 20 mL glass vial.

5.2 Add 10 mL of 1 N HCl. Vortex for 2 minutes.

5.3 Allow the sample to separate for 2 minutes.

5.4 Pipette approximately 1 mL of the supernatant into an appropriately labeled microcentrifuge tube.

5.5 Centrifuge at 6000 rpm for 5 minutes.

5.6 Pipette the supernatant into an appropriately labeled 4 mL glass vial with a Teflon-lined cap.

*Note: Samples will separate into three distinct layers, a solid bottom layer, a slightly cloudy liquid supernatant layer, and a thin white layer on top of the liquid supernatant. Pipette only the center liquid supernatant portion into the 4 mL glass vial.*

- 5.7 Add 3.96 mL of ABRAXIS® Glyphosate Diluent to a clean, appropriately labeled 4 mL glass vial. Add 40 µL of the supernatant (from step 5.6) to the ABRAXIS® Glyphosate Diluent in the vial (1:100 sample dilution). Vortex.
- 5.8 This will then be analyzed as sample, see *Derivatization of Standards, Control, and Samples* in the Test Preparation section of the ABRAXIS® Glyphosate Plate ELISA Kit user's guide.

## 6. Evaluation of Results

The Glyphosate concentration in the samples is determined by multiplying the ELISA results by a factor of 1000. Sample extracts showing a concentration lower than standard 1 (0.075 ppb) should be reported as containing < 0.075 ppm of Glyphosate. Samples showing a higher concentration than standard 5 (4.0 ppb) can be reported as containing > 4ppm of Glyphosate or diluted further and re-analyzed to obtain an accurate quantitative result.

## 7. Performance Data

### *Recovery*

Tofu samples (silken, soft, medium, firm, and extra firm) were spiked with various amounts of Glyphosate, extracted as described above, and then derivatized and assayed using the ABRAXIS® Glyphosate Plate ELISA. Average recovery was 93.5%.7.

## 8. For ordering or technical assistance contact

Gold Standard Diagnostics

795 Horsham Road

Horsham, PA 19044

WEB: [www.abraxiskits.com](http://www.abraxiskits.com)

Phone: (215) 357 3911

Fax: (215) 357 5232

Ordering: [info.abraxis@us.goldstandarddiagnostics.com](mailto:info.abraxis@us.goldstandarddiagnostics.com)

Technical Support: [support.abraxis@us.goldstandarddiagnostics.com](mailto:support.abraxis@us.goldstandarddiagnostics.com)

Date this Technical Bulletin is effective: 05/16/2024

Version: 01