


ABRAXIS® Microcystins 0.3 ppb Strip Test Finished Drinking Water 520046


1. Collect Sample



Collect 1 to 2 mL of sample.

2. Transfer

Squeeze the upper bulb to draw sample into the pipette. The required amount of sample will fit into the stem of the pipette. Any excess will flow into the overflow bulb. Squeeze the upper bulb again to dispense the required amount of sample into conical tube.



3. Shake and Incubate



Close the conical, flip-top tube and shake for 30 seconds.

(Dried reagents will dissolve, turning the sample purple.)




Incubate for 10 minutes.



4. Test


Insert test strip into conical, flip-top tube with arrow pointing down.

Incubate for 10 minutes.



5. Dry

Remove test strip. Lay flat and allow to continue developing for 5 minutes.



6. Interpret



CONTROL LINE TEST LINE INTERPRETATION

NO CONTROL LINE PRESENT NO TEST LINE PRESENT INVALID RESULT

CONTROL LINE PRESENT NO TEST LINE PRESENT ≥3 ppb

CONTROL LINE PRESENT MODERATE INTENSITY BETWEEN 0 AND <3 ppb

TEST LINE PRESENT

Gold Standard Diagnostics 795 Horsham Road Horsham, PA 19044 WEB: www.abraxiskits.com

T (215) 357 3911 F (215) 357 5232 Ordering: info.abraxis@us.goldstandarddiagnostics.com Technical Support: support.abraxis@us.goldstandarddiagnostics.com

Version: 02

Date this User Guide is effective: 11/13/2024